



UNIVERSITÀ
DEGLI STUDI
FIRENZE

FLORE

Repository istituzionale dell'Università degli Studi di Firenze

Point-of-care testing of cholesterol and triglycerides for epidemiologic studies: evaluation of the multicare-in system.

Questa è la Versione finale referata (Post print/Accepted manuscript) della seguente pubblicazione:

Original Citation:

Point-of-care testing of cholesterol and triglycerides for epidemiologic studies: evaluation of the multicare-in system / Stefano Rapi; Cristina Bazzini; Camilla Tozzetti; Valentina Sbolci; Pietro Amedeo Modesti. - In: TRANSLATIONAL RESEARCH. - ISSN 1931-5244. - STAMPA. - 153:(2009), pp. 71-76.
[10.1016/j.trsl.2008.11.010]

Availability:

This version is available at: 2158/347442 since:

Published version:

DOI: 10.1016/j.trsl.2008.11.010

Terms of use:

Open Access

La pubblicazione è resa disponibile sotto le norme e i termini della licenza di deposito, secondo quanto stabilito dalla Policy per l'accesso aperto dell'Università degli Studi di Firenze
(<https://www.sba.unifi.it/upload/policy-oa-2016-1.pdf>)

Publisher copyright claim:

(Article begins on next page)

ORIGINAL ARTICLES

Point-of-care testing of cholesterol and triglycerides for epidemiologic studies: evaluation of the multicare-in system

STEFANO RAPI, CRISTINA BAZZINI, CAMILLA TOZZETTI, VALENTINA SBOLCI,
and PIETRO AMEDEO MODESTI

FLORENCE, ITALY

Cardiovascular disease is the leading cause of death in the world with 80% of cardiovascular events that occur in low- and middle-income countries. Reliable data on the prevalence of risk factors in developing countries can be obtained in door-to-door epidemiologic studies with the use of automatic instruments. This study was performed to assess the sensitivity and specificity of a low-cost and manageable point-of-care testing (POCT) instrument (HPS MultiCare-in, Italy) for cholesterol and triglyceride assays. Plasma blood samples were obtained from consecutive subjects referred to our clinic for diagnostic evaluation. The analyzer currently used in our central laboratory (ADVIA 2400; Siemens, Deerfield, Ill) was used as comparison method. The inter-assay imprecision (expressed as variation coefficient) of the MultiCare POCT system was 4.51% (range, 2.38%–8.54%) and was 3.29% (range, 1.06%–7.45%) for cholesterol and triglycerides systems, respectively. The mean percent bias for capillary samples was $3.5 \pm 4.3\%$ for total cholesterol and $-2.4 \pm 4.9\%$ for triglycerides. The difference in results obtained by nonprofessionals compared with professionals (practicability testing) was $0.28 \pm 7.61\%$ and $1.26 \pm 9.86\%$, respectively (*P* value was nonsignificant for both). Sensitivity and specificity measurements were 95.7% and 61.9% (threshold value of cholesterol 190 mg/dL) and 98% and 93.5% (threshold value of triglycerides 170 mg/dL), respectively. POCT instruments are essential to perform epidemiologic studies while avoiding transportation and storage of biologic material. The characteristics of sensitivity and specificity as well as diagnostic accuracy make the POCT instrument useful for obtaining an accurate stratification of a study population. (*Translational Research* 2009;153:71–76)

Abbreviations: CI = confidence interval; CV = cardiovascular; EDTA = ethylenediaminetetraacetic acid; NCEP = National Cholesterol Education Program; POCT = point-of-care testing; WHO = World Health Organization

From the Central Laboratory of Clinical Biochemistry, Careggi Hospital, and the Department of Critical Care Medicine, University of Florence, Florence, Italy.

Supported by the Ministero dell'Università e della Ricerca, Direzione Generale per le strategie e lo sviluppo dell'internazionalizzazione della ricerca scientifica e tecnologica.

Submitted for publication October 8, 2008; revision submitted November 21, 2008; accepted for publication November 24, 2008.

Reprint requests: Pietro Amedeo Modesti, MD, PhD, Department of Critical Care Medicine and Don Carlo Gnocchi Foundation, University of Florence, Viale Morgagni 85, 50134 Florence, Italy; e-mail: pamodesti@unifi.it.

1931-5244/\$ – see front matter

© 2009 Mosby, Inc. All rights reserved.

doi:10.1016/j.jtrsl.2008.11.010

AT A GLANCE COMMENTARY**Background**

The importance of plasma lipids to stratify global cardiovascular risk is well recognized. However, because of epidemiologic transition, most cardiovascular events occur in developing countries where the prevalence of risk factors is derived from patients referred to clinical centers.

Translational Significance

The possibility of having a low-cost device for cholesterol and triglycerides assay used in epidemiologic door-to-door studies represents a clear example of translational medicine. The device may be used to obtain directly reliable epidemiologic data in the field, which is necessary to plan effective actions to deal with noncommunicable diseases.

Cardiovascular (CV) disease is the leading cause of death worldwide.¹ The World Health Report 1999 estimated that in 1998, 85% of the CV burden developed from low- and middle-income countries.² This contribution is projected to grow, given the consistent population growth rates.^{3,4} Demographic transition and urbanization also contribute to the increased prevalence of risk factors.⁵ Both cholesterol and triglycerides plasma levels are important parameters for assessing risk factors of a population in epidemiologic studies.^{6,7}

Although the developing world bears most of the burden of world CV deaths, there are still no signs of success in halting the CV disease epidemic in developing countries.^{4,8,9} Actions to deal with noncommunicable diseases require consistent data on the prevalence of risk factors to address large-scale prevention programs. This goal can be pursued only with door-to-door epidemiologic studies, which use point-of-care testing (POCT), as suggested by the World Health Organization (WHO).¹⁰ In recent years, semiautomatic POCT instruments that can measure blood levels of cholesterol and triglycerides have been developed. This study was performed to assess the sensitivity and specificity of a new low-cost and manageable POCT instrument (HPS MultiCare-in, Biochemical System International, Arezzo, Italy) for cholesterol and triglyceride assays.

METHODS

Blood samples. The research conforms to the relevant ethical guidelines for human research (Declara-

tion of Helsinki). Informed consent was obtained, and the study was approved by our Institutional Review Board. According to the manufacturer, determinations with the MultiCare system can be performed in ethylenediaminetetraacetic acid (EDTA)-venous blood samples or fresh capillary blood applied directly to the test strip. Blood samples were then collected into K₂EDTA-containing tubes from outpatients (n = 636; age range, 22–69 years) referred to our unit for diagnostic evaluation. Fresh capillary blood samples (n = 66) were obtained immediately before assay in a practicability test.

Device used and quality control. The comparison method involved daily quality controlled with sets of cholesterol and triglyceride solutions at 2 medical decision levels (106 ± 2 and 262 ± 5 mg/dL for cholesterol and 88 ± 2.5 and 173 ± 3.5 mg/dL for triglycerides). In addition, reference instruments were submitted to external quality assurance services (EQAS November 2007 to April 2008; Bio-Rad Laboratory, Hercules, Calif). Bias and imprecision were 0.79 and 1.73 for cholesterol and 2.69 and 2.39 for triglycerides, respectively.

Ten MultiCare devices and 5 different lots of reactive strips for cholesterol (MultiCare, CH) and triglycerides (MultiCare, TGL) were tested during the study. All measurements were performed at room temperature, between 22°C and 28°C. Each MultiCare System device was checked twice per day using the control solution. The MultiCare systems were considered technically reliable if the values obtained were within the control range set up by the manufacturer. Instrument absorbance was reconsidered at the end of the study to confirm instrumental efficiency. The reliability of each lot of reactive strips was tested using reference solutions (expected range of variability set up by the manufacturer for control materials was $\pm 25\%$ for both cholesterol and triglycerides).

Analytical data analysis. Imprecision study. The inter-assay imprecision was calculated by performing 12 runs of the same fresh venous plasma-EDTA sample on the same instrument. Different levels of cholesterol and triglyceride concentrations were studied. The intra-assay imprecision was calculated by performing 4 runs with the same sample on 3 different days. The venous plasma-EDTA used for intra-assay imprecision profiles was stored at 4°C, and hematocrit, cholesterol, and triglycerides levels were tested with reference methods at the beginning and at the end of the study. Intra-assay imprecision was also calculated on control solutions provided by the manufacturer. Imprecision values are expressed as coefficients of variation.




Fig 1. Passing-Bablok comparison of methods for cholesterol ($n = 103$; cholesterol range, 120–370 mg/dL) ($y = 0.9637x + 22.547$; $r = 0.942$) (A). Difference between methods (B).

Accuracy. The reference method was the ADVIA Analyzer (2400; Siemens). According to the manufacturer's instructions, subjects with hematocrit values beyond the 35–50 range were preliminarily excluded. The comparison between methods was performed according to the National Committee for Clinical Laboratory Standards NCCLS EP9¹⁰ using samples with a wide range of cholesterol and triglycerides. Each sample was tested twice with both the MultiCare system and the comparison method. Duplicate results with variability $>15\%$ were retested, and aberrant data were excluded from additional analysis.

Practicability test. In a practicability test, patients self-measured their triglycerides ($n = 34$) or cholesterol ($n = 32$) with the MultiCare system, on capillary blood according to the instruction manual. Immediately afterward, a professional operator conducted a second determination with the same instrument on a second capillary sample. The pairs of results were then compared.

Sensitivity and specificity. The sensitivity and specificity of the MultiCare system (ROC curve data) were ob-




Fig 2. Passing-Bablok regression for cholesterol in practicability test ($n = 32$) ($y = 1.1208x - 20.677$) (A). Differences in results obtained by nonprofessionals compared with professionals (B).

tained by performing measurements with both the POCT device and the comparison method on the same venous EDTA blood sample ($n = 340$). To reproduce the conditions of epidemiologic studies, ROC curves were performed in the total group of subjects independently of hematocrit values.

Statistical analysis. Data shown are mean \pm standard deviation (SD). For statistical analysis, regressions of the method comparisons were calculated according to the method of Passing and Bablok.¹¹ Method differences are presented as Bland-Altman bias plots.¹² All tests were performed using SPSS statistical software (SPSS Inc., Chicago, Ill).

RESULTS

Cholesterol. Imprecision study. The inter-assay imprecision of the MultiCare Cholesterol system was 4.51% (range, 2.38%–8.54%) with cholesterol concentrations ranging between 132 and 368 mg/dL.

The range of intra-assay imprecision on venous-EDTA blood samples was 4.72%–10.17%. The intra-assay




Fig 3. Passing-Bablok comparison of methods for triglycerides ($n = 103$; range of triglycerides, 55–429 mg/dL) ($y = 0.9678x + 7.883$; $r = 0.995$) (A). Differences between methods (B).

imprecision calculated on control material was between 1.27% and 1.72%.

Accuracy. Three of 103 (2.91%) samples were retested for aberrant data (insufficient sample on the strip). When blood samples ($n = 103$; cholesterol range, 120–370 mg/dL) were examined within the framework of the comparison method, the Passing-Bablok regression equation was $y = 0.9637x + 22.547$ (95% confidence interval [CI] for the slope, 0.89–1.031; $r = 0.942$) (Fig 1, A). The mean difference between methods was $3.53 \pm 4.29\%$ (Fig 1, B).

Practicability test. In practicability testing ($n = 32$), the Passing-Bablok regression equation was $y = 1.1208x - 20.677$ (95% CI for the slope, 0.9626–1.279; $r = 0.978$) (Fig 2, A). The difference in results obtained by patients compared with professionals was $0.28\% \pm 7.61\%$ (Fig 2, B).

Sensitivity and specificity. When a cut-off value of 190 mg/dL was considered, the ROC value for MultiCare was 95.7% for sensitivity and 61.9% for specificity.

Triglycerides. Imprecision study. The inter-assay imprecision of the MultiCare triglycerides system was 3.29% (range, 1.06–7.45) with triglyceride concentrations between 79 and 323 mg/dL. The range of




Fig 4. Passing-Bablok regression for triglycerides in practicability test ($n = 34$) ($y = 0.8787x + 16.777$; $r = 0.993$) (A). Differences in results obtained by nonprofessionals compared with professionals (B).

intra-assay imprecision was 2.4%–7.7% on the venous-EDTA blood samples and 1.15%–1.54% on the control solution.

Accuracy. Two samples of 103 (1.95%) were retested for aberrant data (insufficient sample on the strip). The method comparison performed on capillary blood samples ($n = 103$; range of triglycerides 55–429 mg/dL) showed a Passing-Bablok regression equation of $y = 0.9678x + 7.883$ (95% CI for the slope, 0.89–1.031; $r = 0.995$) (Fig 3, A). The mean difference between methods was -2.36 ± 4.95 (Fig 3, B).

Practicability test. In practicability testing ($n = 34$), the Passing-Bablok regression equation was $y = 0.8787x + 16.777$ (95% CI for the slope, 0.9626–1.279; $r = 0.993$) (Fig 4, A). The difference in results obtained by nonprofessionals compared with professionals was 1.26 ± 9.86 (Fig 4, B).

Sensitivity and specificity. When a cut-off value of 170 mg/dL was considered, the ROC values for MultiCare were 98% for sensitivity and 93.5% for specificity.

DISCUSSION

The MultiCare system is easy to learn to use by both professionals and laypersons. A drop of capillary blood

Table I. Analytical performance of POCT device for determination of cholesterol and triglycerides

Assay	Device	Inter-assay Imprecision (%)	Bias (%)	Year (reference)
Cholesterol	Cholestech LDX	4.0	2.1	1998 ¹³
	Cholestech LDX	3.0	—	2007 ¹⁴
	Cholestech LDX	—	−0.6 ± 6.9	2002 ¹⁵
	CardioChek PA	4.4	—	2007 ¹⁴
	Accumeter	5.3	−1.0	1998 ¹³
	Accutrend	2.5	−5.6–16.6	2000 ¹⁶
	Accutrend	<5	−3.2–2.5	1995 ¹⁷
	MultiCare-in	4.5	3.5 ± 4.3	
NCEP goals		<3	<3	1988 ¹⁸
Triglycerides	Cholestech LDX	2.6	—	2007 ¹⁴
	Cholestech LDX	—	22 mg/dL	2006 ¹⁹
	Cholestech LDX	—	−19 ± 9.4	2002 ¹⁵
	CardioChek PA	4.8	—	2007 ¹⁴
	Accutrend	1.4–6.1	0.9 ± 12.9	2000 ²⁰
	MultiCare-in	3.3	−2.4 ± 4.9	
NCEP goals		<5	<5	1988 ¹⁸

is applied to the test strip, and results are obtained after 3 min for cholesterol and 2 min for triglyceride assays. The necessary minimum volume is 10 μ L. The handling procedure is much the same as those used with most blood glucose meters, but certain precautions should nevertheless be taken. The user's hands must be washed with warm water before measurement to avoid possible contamination with glycerol from hand lotions, soaps, or disinfectants. In addition, the blood drop must be obtained strictly according to the instructions. When device quality control was correct and these simple precautions were followed, laypersons achieved results comparable with those achieved by professionals.

The MultiCare systems are pocket-sized reflectance photometers, in which the intensity of the color developed from a chromogen reaction being proportional to the concentration of the cholesterol or triglycerides in the blood. The results of the MultiCare method compared with the reference method demonstrated good agreement between the 2 methods, with a mean difference of 3.5% and −2.3% for cholesterol and triglycerides, respectively.

The availability of POCT lipid monitors has increased in recent years (see Table I).^{13–20} Any POCT must be validated for bias and imprecision to ensure that appropriate medical decisions and population screenings are made.^{21–23} Bias (inaccuracy) is defined as disagreement between the monitor and the standardized laboratory. The National Cholesterol Education Program (NCEP) in the United States recommended bias goals of 3% and 5% for cholesterol and triglycerides, respectively.^{18,24} Imprecision refers

to the reproducibility of a test result. Imprecision goals of 3% and 5% are the desirable specification for total cholesterol and triglyceride laboratory methods, respectively, according to the NCEP.^{18,24} Overall, the analytical goals for POCT should be equivalent to those used for laboratories to ensure that POCT use does not compromise standards of patient care and clinical decision making. POCT methods are not designed to replace lipid determinations in professional laboratories. In addition, it is important to acknowledge that particularly in rural areas, the POCT environment is different from the laboratory setting. Therefore, the most recent analytical goals recommended for POCT instruments used for diabetes management in the public health setting in Australia considered a minimum imprecision goal of 5% for cholesterol and 7.5% for triglycerides.²⁵

Cholesterol, triglyceride, and blood glucose assays are the 3 laboratory investigations required by the WHO Stepwise approach to Surveillance, which is a simple, standardized method for collecting, analyzing, and disseminating data in WHO member countries.²⁶ Other currently available POCT devices can measure high-density lipoprotein and low-density lipoprotein directly, which avoids the need for testing while the patient is fasting. However, fasting is usually required in epidemiologic studies to obtain reliable data on diabetes prevalence. It must be noted that multiple operators and instruments involved in our study simulated a true picture of the usual clinical setting. The good results of the practicability test as well as the small amount of blood needed to perform measurements make the MultiCare instrument useful in epidemiologic

studies. The investigator may test the subject at home thus obtaining reliable data in the field. This aspect is crucial in developing countries in which the transfer of biologic material from remote areas to a clinical center may introduce a significant source of variability because of unpredictable problems in sample storage and refrigeration. Some potential advantages of this advice for developing country screening are that it or the test strips do not require refrigeration, and that is small, uses long-life batteries, and it is not costly in terms of the device or the strips. In particular, the cost of the tested device (Multicare-in, \$48) is less compared with other available instruments.

The instrument might also be useful for monitoring responses to therapy in a single patient. The easy operation of the system enables patients to monitor their own blood lipids in response to lifestyle changes, such as modified eating and exercise patterns or other therapeutic means adopted.

REFERENCES

1. Yusuf S, Reddy S, Ounpuu S, Anand S. Global burden of cardiovascular diseases: part I: general considerations, the epidemiologic transition, risk factors, and impact of urbanization. *Circulation* 2001;104:2746–53.
2. The World Health Report. Making a difference. Geneva, Switzerland: World Health Organization, 1999.
3. Mathers CD, Loncar D. Projections of global mortality and burden of disease from 2002 to 2030. *PLoS Med* 2006;3:e442.
4. Reddy KS. Cardiovascular disease in non-Western countries. *N Engl J Med* 2004;350:2438–40.
5. World Health Organization. The world health report 2002: reducing risks, promoting healthy lifestyles. Geneva, Switzerland: World Health Organization, 2003.
6. Grundy SM, Cleeman JI, Merz CN, et al. National Heart, Lung, and Blood Institute; American College of Cardiology Foundation; American Heart Association. Implications of recent clinical trials for the National Cholesterol Education Program Adult Treatment Panel III guidelines. *Circulation* 2004;110:227–39.
7. Grundy SM, Brewer HB Jr., Cleeman JI, Smith SC Jr., Lenfant C. American Heart Association; National Heart, Lung, and Blood Institute. Definition of metabolic syndrome: report of the National Heart, Lung, and Blood Institute/American Heart Association conference on scientific issues related to definition. *Circulation* 2004;109:433–8.
8. Joshi R, Jan S, Wu Y, MacMahon S. Global inequalities in access to cardiovascular health care: our greatest challenge. *J Am Coll Cardiol* 2008;52:1817–25.
9. Murray CJ, Lopez AD. Alternative projections of mortality and disability by cause 1990–2020: Global Burden of Disease Study. *Lancet* 1997;349:1498–504.
10. National Committee for Clinical Laboratory Standards. Method comparison and bias estimation using patient samples, tentative guideline. NCCLS publication EP9-A. Villanova, Pa: National Committee for Clinical Laboratory Standards, 1995. 1–36.
11. Passing H, Bablok W. A new biometrical procedure for testing the equality of measurements from two different analytical methods. *J Clin Chem Clin Biochem* 1983;21:709–20.
12. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986;8:307–10.
13. Volles DF, McKenney JM, Miller WG, Ruffen D, Zhang D. Analytic and clinical performance of two compact cholesterol-testing devices. *Pharmacotherapy* 1998;18:184–92.
14. Shephard MD, Mazzachi BC, Shephard AK. Comparative performance of two point-of-care analysers for lipid testing. *Clin Lab* 2007;53:561–6.
15. Stein JH, Carlsson CM, Papcke-Benson K, Einerson JA, McBride PE, Wiebe DA. Inaccuracy of lipid measurements with the portable Cholestech L.D.X analyzer in patients with hypercholesterolemia. *Clin Chem* 2002;48:284–90.
16. du Plessis M, Ubbink JB, Vermaak WJ. Analytical quality of near-patient blood cholesterol and glucose determinations. *Clin Chem* 2000;46:1085–90.
17. Gottschling HD, Reuter W, Ronquist G, Steinmetz A, Hattemer A. Multicentre evaluation of a non-wipe system for the rapid determination of total cholesterol in capillary blood, Accutrend Cholesterol on Accutrend GC. *Eur J Clin Chem Clin Biochem* 1995;33:373–81.
18. Laboratory Standardisation Panel of the National Cholesterol Education Program. Current status of blood cholesterol measurement in clinical laboratories in the United States: a report from the Laboratory Standardisation Panel of the National Cholesterol Education Program. *Clin Chem* 1988;34:193–201.
19. Carey M, Markham C, Gaffney P, Boran C, Maher V. Validation of a point of care lipid analyser using a hospital based reference laboratory. *Ir J Med Sci* 2006;175:30–5.
20. Luley C, Ronquist G, Reuter W, et al. Point-of-care testing of triglycerides: evaluation of the Accutrend triglycerides system. *Clin Chem* 2000;46:287–91.
21. Gregory LC, Duh SH, Christenson RH. Eight compact analysis systems evaluated for measuring total cholesterol. *Clin Chem* 1994;40:579–85.
22. Krouwer JS. Setting performance goals and evaluating total analytical error for diagnostic assays. *Clin Chem* 2002;48:919–27.
23. Taylor JR, Lopez LM. Cholesterol: point-of-care testing. *Ann Pharmacother* 2004;38:1252–7.
24. Warnick GR, Myers GL, Cooper GR, Rifai N. Impact of the third cholesterol report from the adult treatment panel of the national cholesterol education program on the clinical laboratory. *Clin Chem* 2002;48:11–7.
25. Shephard MDS. Analytical goals for point-of-care testing used for diabetes management in Australian health care settings outside the laboratory. *Point of Care* 2006;5:177–85.
26. World Health Organization. STEPwise approach to surveillance (STEPS). Available at: <http://www.who.int/chp/steps/en/>. Accessed August 2008.